























positioning  
patients for life.®

# SYMBOLS GLOSSARY

SYMBOL	SYMBOL TITLE	DESCRIPTION	STANDARD TITLE & DESIGNATION NUMBER	REFERENCE NUMBER
	Date of Manufacture	Indicates the date when the medical device was manufactured.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.3
	Manufacturer	Indicates the medical device manufacturer, as defined in EU Directives 90/385/EEC, 93/42/EEC and 98/79/EC.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.1
	Authorized representative in the European community	Indicates the Authorized Representative in the European Community.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.2
	Conformité Européene (European Conformity)	CE Marking on a product is a manufacturer's declaration that the product complies with the essential requirements of the relevant European health, safety and environmental protection legislation.	Directive 93/68/EEC.	N/A
	Reorder Number	Indicates the manufacturer's catalogue number so that the medical device can be identified.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.6
	Serial Number	Indicates the manufacturer's serial number so that a specific medical device can be identified.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.7
	Batch Code	Indicates the manufacturer's batch code so that the batch or lot can be identified.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.1.5
	Refer to instruction manual/booklet	To signify that the instruction manual/booklet must be read.	ISO 7010—Graphical Symbols—Safety Colors and Safety Signs—Registered Safety Signs	ISO 7010-M002
	MR Safe	An item that poses no known hazards resulting from exposure to any MR environment.	ASTM F2503 - 13 Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment	7.3.1

SYMBOL	SYMBOL TITLE	DESCRIPTION	STANDARD TITLE & DESIGNATION NUMBER	REFERENCE NUMBER
	MR Conditional	An item with demonstrated safety in the MR environment within defined conditions.	ASTM F2503 – 13 Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment	7.3.2
	MR Unsafe	An item which poses unacceptable risks to the patient, medical staff or other persons within the MR environment.	ASTM F2503 – 13 Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment	7.3.3
	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.4.4
	General warning sign	To signify a general warning	ISO 7010—Graphical Symbols—Safety Colors and Safety Signs—Registered Safety Signs	ISO 7010-W001
	Warning; Electricity	To warn of electricity	ISO 7010—Graphical Symbols—Safety Colors and Safety Signs—Registered Safety Signs	ISO 7010-W012
	Do Not Reuse	Indicates a medical device that is intended for one use or for use on a single patient during a single procedure.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.4.2
	Type B Applied Part	On medical equipment. To identify a type B applied part complying with IEC 60601-1.	IEC/TR 60878 Graphical symbols for electrical equipment in medical practice.	5840
	Keep Dry	Indicates a medical device that needs to be protected from moisture.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.3.4
	Fragile, handle with care	Indicates a medical device that can be broken or damaged if not handled carefully.	ISO & ANSI/AAMI/ISO 15223-1 Medical devices—Symbols to be used with medical device labels—General requirements.	5.3.1
	Quantity	Indicates quantity of medical devices contained within the packaging.	N/A	N/A
	Purchase Order	Indicates the Customer Purchase Order Number for the purchase of the medical device contained within the packaging.	N/A	N/A